



Calibration Certificate

No. 04920156/001

Model R0110
Serial Number 04920156
Number of pages 1

Process of Calibration Direct comparison

Conditions of Measurement

Temperature	23 °C	± 3 °C
Relative Humidity	40 %	± 20 %

Calibration Standards (valid to)

Thermometer, PE105, 8326/26, (29.8.2004)

All standards are traceable to ČMI (Czech Metrology Institute) or in terms of relative humidity to DKD Testo (Deutscher Kalibrierdienst).

Measuring Results

Channel 1-Temperature Intern

Standard	Device	Uncertainty	Used Standard	Note
22.82 °C	22.6 °C	0.21 °C	PE105	

Measuring Uncertainty

The expanded uncertainty of measurement corresponding to the measurement results is started as the standard uncertainty of measurement multiplied by the coverage factor $k=2$. Usually the values is located in the corresponding interval with probability of approximately 95%. This was determined in accordance with EA4/02.

Date of Calibration

14.1.2004

Calibrated By

.....
Čestmír Zeman

Has Approved

.....
Milan Jurajda